



Diversey Europe BV
 Maarssebroeksedijk 2
 3542 DN Utrecht
 The Netherlands
 Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 31/7/12

Product Name	CLEARLENs IPA VH1
Product Code	7513400
Batch Number	FMP12191 32412
Production Date	07/2012
Expiration Date	07/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear colourless liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	<i>[Signature]</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

CERTIFICATE OF IRRADIATION

Number: **B**

0539729 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

P018861

Cust. Ref: 19/07/12

Date Rec'd: 24/07/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212070580	JD IPA BOTTLES 25-45kGy	705	B/No.W032412 FMP12191. <i>Q-03/6/13</i> ROUTINE DOSE 37.3, 37.1, 38.1, 39.6 kGy
FLE0010999 B212070580	SAMPLES	5	CUSTOMER BATCH/CODE No.s AS ORDER No.P018861 DATED 18/7/12
Despatch Text: 13PLTS		Total	710
			Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *T. Coffey*

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 01 Aug 2012
Date Tested: 07 Aug 2012
Date Test Completed: 07 Aug 2012
Purchase Order: PO18901

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0005988/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 01/08/2012
Test Article: Clearklens IPA Block 1
Sample Code: 3607
Batch Ref: FMP12191 32412
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	61 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Suz Wood

Date: 08 Aug 2012 13:37:36

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0005988

Certificate of Analysis - OSMM

Print Number: P0004586

Page 1 of 1



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 01 Aug 2012
Date Tested: 09 Aug 2012
Date Test Completed: 23 Aug 2012
Purchase Order: PO18901

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0005988/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 01/08/2012
Test Article: Clearklens IPA Block 1
Sample Code: W032412, 3607
Batch Ref: FMP12191 32412
Qty Received: 20 x 900mL Bottles + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 28 Aug 2012 11:47:52

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology