



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 14/8/12

| | |
|-----------------|-------------------|
| Product Name | CLEARLENS IPA VH1 |
| Product Code | 7513400 |
| Batch Number | FMP12206 32487 |
| Production Date | 07/2012 |
| Expiration Date | 07/2014 |

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

| Test | Test Method | Limits | | Results |
|-------------------------|-------------|-------------------------|-------|-------------------------|
| | | Lower | Upper | |
| Appearance | Visual | Clear Colourless Liquid | | Clear colourless liquid |
| Specific Gravity (20°C) | JDM004 | 0.872 | 0.883 | 0.877 |
| | | | | |
| | | | | |

| | | |
|---|----------|---------------------------|
| On behalf of Diversey site Quality Manager | Name : | |
| | Position | Quality Control Inspector |

This document being issued electronically does not bear a signature

| | | |
|-------------|--------------|--|
| COA 7513400 | Version : 03 | Date of issuing : November 28 th 2011 |
|-------------|--------------|--|

CERTIFICATE OF IRRADIATION Number: **B** 0541476 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster

Synergy Health plc
 Roydsdale Way · Euoway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com



LA1 4XS
 Cust. Ref: P018919
 Date Rec'd 2/08/12
 Date: 6/08/12

| ITEM CODE SYNERGY HEALTH BATCH | ITEM SPECIFICATION | QTY | ADDITIONAL DETAILS |
|-----------------------------------|-------------------------|-----|---|
| B2FLE0010081 B212080117 | JD IPA BOTTLES 25-45kGy | 712 | B/No. W032487 FMP 1220b, <i>Don 031613</i> ROUTINE DOSE 38.7, 37.4, 37.8, 38.9 kGy |
| B2FLE0010999 B212080117 | SAMPLES | 5 | CUSTOMER BATCH/CODE No.s AS ORDER No. P018919 DATED 1/8/12 |
| Despatch Text: 14PLTS | | | |
| Total | | 717 | Last Page 1 of 1 |

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *T. Coffey*

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 13 Aug 2012
Date Tested: 15 Aug 2012
Date Test Completed: 15 Aug 2012
Purchase Order: -

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0006170/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 13/08/2012
Test Article: Clearklens IPA Block 1
Sample Code: 3652
Batch Ref: FMP 12206 32487
Qty Received: 1 x 900mL Bottle

| Test | Method Item | Result |
|---|-------------|------------|
| Bacterial Endotoxin by Kinetic QCL Test | MM110/00 | <0.2 EU/mL |
| Spiked Recovery | MM110/01 | 101 % |
| Dilution Tested | MM110/04 | 1/40 |
| Endotoxin Spike Level | MM110/02 | 0.5 EU/mL |
| Product Standard Data Sheet | FG047/psd | FM02 |

Comments

2 Boxes received on the 13/08/12, and 2 boxes received on the 14/08/12
Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature, Their name is shown below.

S. L. Wood

Date: 16 Aug 2012 12:29:56

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 14 Aug 2012
Date Tested: 16 Aug 2012
Date Test Completed: 30 Aug 2012
Purchase Order: P018977

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0006170/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 14/08/2012
Test Article: Clearklens IPA Block 1
Sample Code: 3652
Batch Ref: FMP 12206 32487
Qty Received: 20 x 900mL Bottles + 3 spare

| Test | Method Item | Result |
|--|-------------|------------------------|
| Sterility Test by Membrane Filtration (Steritest) Method | MM107/00 | Pass |
| Growth in Tryptone Soya Broth at 20-25°C after 14 days | MM107/01 | No growth in one broth |
| Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days | MM107/02 | No growth in one broth |
| Volume Tested | MM107/05 | 20 x 50 mL |
| Product Standard Data Sheet | FG047/psd | FM03 |

Comments

2 Boxes received on the 13/08/12, and 2 boxes received on the 14/08/12
Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 03 Sep 2012 09:35:48

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0006170

Certificate of Analysis - OSMM

Print Number: P0005149

Page 1 of 1