



Diversey Europe BV
 Maarssebroeksedijk 2
 3542 DN Utrecht
 The Netherlands
 Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 16/10/12

Product Name	CLEARLENs IPA VH1
Product Code	7513400
Batch Number	FMP12264 32539
Production Date	09/2012
Expiration Date	09/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear colourless liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	<i>C. ASCOR</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

CERTIFICATE OF IRRADIATION Number: **B** 0547845 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

Synergy Health plc
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@synergyhealthplc.com
Website: www.synergyhealthplc.com



LA1 4XS
P019203
27/09/12
8/10/12

Cust. Ref:
Date Rec'd
Date:

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212100038	JD IPA BOTTLES 25-45kGy	696	B/No.W032539 FMP 12264 32539, (EC) 23.10.12. ROUTINE DOSE 38.4, 37.0, 37.0, 39.7 kGy's
B2FLE0010999 B212100038	SAMPLES	5	CUSTOMER BATCH/CODE No.s AS ORDER No.P019203 DATED 26/9/12
Despatch Text: 13PLTS			
Total		701	Last Page 1 of 1

All in accordance with current
Technical Agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *[Handwritten Signature]*

For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 Oct 2012
Date Tested: 17 Oct 2012
Date Test Completed: 31 Oct 2012
Purchase Order: 19305

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007195/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 15/10/2012
Test Article: Clearklens IPA Block 1
Sample Code: 3804
Batch Ref: FMP12264 32539
Qty Received: 20 x 900mL Bottles + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

S.L. Wood

Date: 05 Nov 2012 11:07:47

Mrs S.L. Wood, BSc (Hons), CBIOL, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007195

Certificate of Analysis - OSMM

Print Number: P0006673

Page 1 of 1



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 Oct 2012
Date Tested: 18 Oct 2012
Date Test Completed: 18 Oct 2012
Purchase Order: 19305

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0007195/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 15/10/2012
Test Article: Claerklen IPA Block 1
Sample Code: 3804
Batch Ref: FMP12264 32539
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	96 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 22 Oct 2012 16:02:18

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0007195

Certificate of Analysis - OSMM

Print Number: P0006265

Page 1 of 1