



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W026350
PRODUCTION DATE	08/2009
EXPIRY DATE	08/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless Liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.884

Date: 26/08/09

QA Manager: [Signature]

QA: E. Gavanis

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014389
Date Rec'd 13/08/09
Date: 20/08/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209080347	JD DE 900ML BOTTLES 25-45kGy	705	B/No.W026350 - CUSTOMER B/No.933 ACTUAL....36.1 - 37.9....kGy DOSE
[FLE0010999 B209080347	SAMPLES	2	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014389 DATED 12/8/09
Total		707	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 27-Aug-09
Date Tested : 02-Sep-09
Date Completed : 02-Sep-09
Purchase Order No. : P014475

Test Article : Clearklens DE Block 1

Sample Code : Works Order No. 26350

Batch Reference : 933

Laboratory Reference : LR-00884096

Article Quantity : 1 Bottles

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01


This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.

Test carried out according to current Pharmacopoeias.

**Mr K.A. Barker C.Biol., M.I. Biol.,
Business Manager**

Date 11 Sep 2009

K.A. Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology



Sample Booking Number: CN\00139998
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 27-Aug-09
Date Tested : 03-Sep-09
Date Completed : 17-Sep-09
Purchase Order No. : P014475

Test Article : Clearklens DE Block 1
Sample Code : Works Order No. 26350
Batch Reference : 933
Laboratory Reference : LR-00884095 Article Quantity : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 574 045
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Mrs S L Wood, BSc., CBiol., MIBiol
Laboratory Manager - Pharmaceutical Microbiology

Date 24 Sep 2009

GLP

Compliant
Laboratory

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00139998

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited