



Diversey Europe BV

Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427

### CERTIFICATE OF ANALYSIS

Date: 20/2/12

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	W031297
Production Date	02/2012
Expiration Date	02/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow	Clear slightly yellow
pH 100% (20°C)	JDM001	7.5	8.5	7.6
Specific Gravity (20°C)	JDM004	0.993	1.003	1.001

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

*This document being issued electronically does not bear a signature*

COA 7513387	Version : 03	Date of issuing : November 28 <sup>th</sup> 2011
-------------	--------------	--

FLE001

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster



**Synergy Health plc**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@synergyhealthplc.com  
Website: www.synergyhealthplc.com

LA1 4XS  
Cust. Ref: P018143  
Date Rec'd 9/02/12  
Date: 13/02/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B212020390	DIVERSEY TEGO 2000SC 50ml 25-42kGy	102	B/No.W031297 - CUSTOMER B/No.387841  ROUTINE DOSE 33.6, 33.4 kGy
B2FLE0010999 B212020390	SAMPLES	1	CUSTOMER BATCH/CODE No.5 AS ORDER No.P018143 DATED 8/2/12
Despatch Text:  2PLTS			
Total		103	Last Page 1 of 1

All in accordance with current Technical Agreement

---

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Cooper*

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 20 Feb 2012  
**Date Tested:** 23 Feb 2012  
**Date Test Completed:** 16 Mar 2012  
**Purchase Order:** 18195

### CERTIFICATE OF ANALYSIS

**Laboratory Reference Number:** 0003645/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: PH Eur and USP  
**Date Received:** 20/02/2012  
**Test Article:** Clearklens Tego 2000sc VH25  
**Sample Code:** 387841  
**Batch Ref:** Works Order No. 31297  
**Qty Received:** 20 x 50mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

### Comments

As Steritest Sterility Test

Approval is provided by Electronic Signature. Their name is shown below.

Date: 20 Mar 2012 08:26:25

Mrs P. Pham-Lengoc, BSc, MSB  
Laboratory Manager - Pharmaceutical Microbiology