

CERTIFICATE OF ANALYSIS

Date issue: 17-Aug-20


Product Name	ClearKlens Tego 2000SC VH25
Product Code	100868202
Batch Number	FMP19124 47409
Production date	04.05.2019
Expiration Date	04.05.2022

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Conformity
Appearance	Visual	Liquid Clear Slightly Yellow	N/A	Liquid Clear Slightly Yellow
Specific Gravity (20°C)	DM 004	0.993 – 1.003	g/cm ³	1.002
pH (neat Solution)	DM 001	7.5 – 8.5	N/A	7.6

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 45.0 (kGy)	Conform	2173-3390A
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	Pass	0049051/2

	Date of released On behalf of Diversey's Manufacturing location: 05/08/2019	by : Juliana Hamberland Position: 3PA Quality Manger
---	---	---

The analysis results above could change over time ant in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the MSDS.

End of certificate of analysis