



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2001 SC VH24
PRODUCT CODE	7513402
BATCH NUMBER	W023127
PRODUCTION DATE	11 / 2007
EXPIRY DATE	11 / 2009

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless to Yellow Liquid	Yellowish liquid
Odour		Typical	
pH (neat)	GM 0130 01	8.5 - 9.5	9.1
Density (20°C)	GM 0110 01	1.060 - 1.070	1.064
Refractive Index (20°C)	GM 0120 01	1.374 - 1.384	1.384

Date:..... 13/11/07

QA Manager:..... *[Signature]*

QA:..... *[Signature]*

JD Approval:..... *[Signature]*

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

E001,

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron plc

Roydsdale Way · Euroway Trading Estate
 Bradford
 West Yorkshire · BD4 6SE
 Telephone: +044 (0)1274 686011
 Facsimile: +044 (0)1274 686061
 Email: bradford@isotron.co.uk
 Website: www.isotron.co.uk

LAI 4XS

st. Ref: PD11969
 te Rec'd 1/11/07
 te: 6/11/07

ITEM CODE ISOTRON BRANCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
FLE0010091 07110085	DIVERSEY TEG02001 10X5x100 25-38kGy	43	36 X B/No.W023127 - CUSTOMER B/No. MK07202413 3 X B/No.W023128 - CUSTOMER B/No. 235300 4 X B/No.W023129 - CUSTOMER B/No. MK0752496 ACTUAL... 32.9 - 33.4 ... kGy DOSE
FLE0010999 07110085	SAMPLES	3	CUSTOMERS BATCH/CODE No.s AS ORDER No.P011969 DATED 31/10/07
FLE001MIN	MINIMUM CHARGE UPLIFT		<p style="text-align: right;"><i>C. Drake</i> 06/11/07</p>
Total		46	Last Page 1 of 1

All in accordance with current technical agreement ✓

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drake*

For and on behalf of ISOTRON plc