




CERTIFICATE OF ANALYSIS


PRODUCT NAME	CLEARLENDS DE
PRODUCT CODE	6069643
BATCH NUMBER	W024288
PRODUCTION DATE	07 / 2008
EXPIRY DATE	07 / 2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.880

Date:..... 30/07/08

QA.Manager:..... 

QA:..... 

JD Approval:..... 

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

CERTIFICATE OF IRRADIATION Number: **B** 0377121 00

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Page 1

Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P012882-883-884-885
 Date Rec'd 17/07/08
 Date: 22/07/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B208070431	JD DE 900ML BOTTLES 25-45kGy	348	B/No.W024288 CUSTOMER B/No.29373 ACTUAL...37.3 - 38.4...kGy DOSE
B. LE0010999 B208070431	SAMPLES	4	CUSTOMERS BATCH/CODE No.s AS ORDER No.P012882 DATED 16/7/08
Total		352	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 28-Jul-08
Date Tested : 31-Jul-08
Date Completed : 31-Jul-08
Purchase Order No. : PO12922

Test Article : Clearklens DE

Sample Code :

Batch Reference : WO24288

Laboratory Reference : LR-00776127

Article Quantity : 1 x900mL

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test	Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	* MM110	<0.25 EU/mL
Dilution	* MM110/01	1/50
Spiked Recoveries	* MM110/02	Valid 91%
Endotoxin Spike Level	* MM110/09	0.5 EU/mL

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM01

GLP
Compliant
Laboratory

MB

Date 8 Aug 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00124385
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 28-Jul-08
Date Tested : 07-Aug-08
Date Completed : 21-Aug-08
Purchase Order No. : PO12922

Test Article : Clearklens DE
Sample Code :
Batch Reference : WO24288
Laboratory Reference : LR-00776126 **Article Quantity** : 20 x900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 226020

Volume Tested: 20x50mL

Test carried out according to current Pharmacopoeias.

(3 bottles were found to be leaking when received, client contacted about how to proceed with testing, proceed with testing.)

GLP

Compliant
Laboratory

WKB

Date 1 Sep 2008

K.A.Barker. C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00124385
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited