

JohnsonDiversey
Clean is just the beginning



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	W024360
PRODUCTION DATE	07/2008
EXPIRY DATE	07/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	Clear Colourless Liquid
Density (20°C)	GM 0110 01	0.876 - 0.896	0.877

Date: 21/07/08

QA Manager: [Signature]

QA: [Signature]

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

Issue no. 4
Issued by W. Ashton
Issue date 07.04.2006

CERTIFICATE OF IRRADIATION Number: **B** 0377914 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P012914- P012919
 Date Rec'd 24/07/08
 Date: 30/07/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B208070647	JD DE 900ML BOTTLES 25-45kGy	348	FMP B/No. W024360 CUST B/No. F/C 29373 ACTUAL...37.1 - 39.8...kGy DOSE
B2FLE0010999 B208070647	SAMPLES	3	
Total		351	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 04-Aug-08
Date Tested : 21-Aug-08
Date Completed : 21-Aug-08
Purchase Order No. : PO12963

Test Article : ClearKlens DE
Sample Code : WO24360
Batch Reference : 29373
Laboratory Reference : LR-00777796 **Article Quantity** : 1
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	72% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.
Product Standard Data Sheet: FM01



WAB

Date 1 Sep 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00124612
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 04-Aug-08
Date Tested : 06-Aug-08
Date Completed : 20-Aug-08
Purchase Order No. : PO12963

Test Article : ClearKlens DE
Sample Code : WO24360
Batch Reference : 29373
Laboratory Reference : LR-00777795 Article Quantity : 20
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 226022
Volume Tested: 20x50mL
Test carried out according to current Pharmacopoeias.



WAB

Date 28 Aug 2008

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00124612
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited