



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W024966
PRODUCTION DATE	10/2008
EXPIRY DATE	10/2010

TEST	METHOD	LIMITS	RESULTS
Appearance		Clear Colourless Liquid	clear colourless liquid.
Density (20°C)	GM 0110 01	0.865 - 0.885	0.877

Date: 05/11/08

QA Manager: [Signature]

QA: E. G. [Signature]

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013312
 Date Rec'd 23/10/08
 Date: 31/10/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010081 B208100659	JD IPA BOTTLES 25-45kGy	511	B/No.W024966 ACTUAL....38.0 - 40.8....kGy DOSE
BZFLE0010999 B208100659	SAMPLES	10	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013312 DATED 23/10/08
Total		521	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 07-Nov-08
Date Tested : 11-Nov-08
Date Completed : 11-Nov-08
Purchase Order No. : PO13366


Test Article : Clearlens IPA
Sample Code : Works Order No. 24966
Batch Reference : 138
Laboratory Reference : LR-00804150 **Article Quantity** : 2 Bottles
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	136% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

Mrs S L Wood, BSc., CBiol., MBIol.
Laboratory Manager - Pharmaceutical Microbiology

Date 19 Nov 2008



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00128455
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS


Date Received : 07-Nov-08
Date Tested : 13-Nov-08
Date Completed : 27-Nov-08
Purchase Order No. : PO13366

Test Article : Clearklens IPA
Sample Code : Works Order No. 24966
Batch Reference : 138
Laboratory Reference : LR-00804149 **Article Quantity** : 20 Bottles
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 306023
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.

Date 09 Dec 2008



K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology