





CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7514372
BATCH NUMBER	100 27317
PRODUCTION DATE	01/2010
EXPIRY DATE	01/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.865 – 0.885	0.878

QA Manager:  Date: 29/01/2010

QA:  Date: 29-01-10

JD Approval:  Date: 29/01/2010

Date above relates to the date of signing of the document by each responsible person.

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014978
Date Rec'd 19/01/10
Date: 25/01/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B210010371	JD IPA BOTTLES 25-45kGy	705	CUSTOMERS BATCH/CODE NO'S AS P014978 DATED 18.01.10 FMP BATCH NO: W027317 CUSTOMERS BATCH NO: 1494 ACTUAL DOSE....37.3 - 38.4....kGy
B2FLE0010999 B210010371	SAMPLES	6	
	Despatch Text: 14PLTS		
Total		711	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 01-Feb-10
Date Tested : 03-Feb-10
Date Completed : 03-Feb-10
Purchase Order No. : 15059

Test Article : Clearklens IPA Block 2

Sample Code :

Batch Reference : WO27317

Laboratory Reference : LR-00927584

Article Quantity : 1 x 900mL

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	Valid 56%
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

Suzanne Wood

Date 05 Feb 2012

Mrs S.L. Wood, B.Sc. (Hons), C.Biol., MSB.,
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00146125

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 01-Feb-10
Date Tested : 04-Feb-10
Date Completed : 18-Feb-10
Purchase Order No. : 15059

Test Article : Clearklens IPA Block 2

Sample Code :

Batch Reference : WO27317

Laboratory Reference : LR-00927583

Article Quantity : 20 x 900mL

Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 745039

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

GLP

Compliant
Laboratory

Suz Wood

Date *22 Feb 2010*

Mrs S.L. Wood, B.Sc. (Hons), C.Biol., MSB.,
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN\00146125
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited