



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W027459
PRODUCTION DATE	02/2010
EXPIRY DATE	02/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 – 8.5	7.5
Density (20°C)	GM 0110 01	0.993 – 1.003	1.003

Date:.....12/02/10.....

QA Manager:.....*[Signature]*.....

QA:.....*[Signature]*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015075
 Date Rec'd 4/02/10
 Date: 9/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210020112	DIVERSEY TEGO 2000SC 50ml 25-42kGy	99	2 X B/No.W027563 CUSTOMER B/No.E568959 97 X B/No.W027459 CUSTOMER B/No.319883 ACTUAL DOSE....32.2 - 34.2....kGy
B2FLE0010999 B210020112	SAMPLES	4	2 X B/No.W027563 2 X B/No.W027459 CUSTOMERS BATCH/CODE No.s AS ORDER No.P015075 DATED 3/2/10
Despatch Text:			
3PLTS			
Total		103	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Goffey*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 17-Feb-10
Date Tested : 24-Feb-10
Date Completed : 10-Mar-10
Purchase Order No. : P015119

Test Article : ClearKlens Tego 2000 SC VH25 Block 1
Sample Code :
Batch Reference : W027459
Laboratory Reference : LR-00932993 **Article Quantity** : 20x50mL +25
Test Required : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 800001
Volume Tested: 20x50mL
Test carried out according to current Pharmacopoeias.

Mr K A Barker, CBiol., MSB
Business Manager

Date 16 Mar 2010



Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Sample Booking Number: CN00146884
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited