

CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W/O 27563
PRODUCTION DATE	02/2010
EXPIRY DATE	02/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 – 8.5	8.0
Density (20°C)	GM 0110 01	0.993 – 1.003	1.002

Date:.....12/02/10.....
 QA Manager:.....
 QA:.....
 JD Approval:.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015075
 Date Rec'd 4/02/10
 Date: 9/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210020112	DIVERSEY TEGO 2000SC 50ml 25-42kGy	99	2 X B/No. W027563 CUSTOMER B/No. ES68959 97 X B/No. W027459 CUSTOMER B/No. 319883 ACTUAL DOSE....32.2 - 34.2....kGy
B2FLE0010999 B210020112	SAMPLES	4	2 X B/No. W027563 2 X B/No. W027459 CUSTOMERS BATCH/CODE No.s AS ORDER No. P015075 DATED 3/2/10
Total		103	Last Page 1 of 1

Despatch Text:

3PLTS

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffey*

For and on behalf of ISOTRON LIMITED