



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 28/5/12

Product Name	CLEARKLENS IPA VH1
Product Code	7513400
Batch Number	WJ32134
Production Date	05/2012
Expiration Date	05/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Colourless	Clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.872	0.883	0.877

On behalf of Diversey site Quality Manager	Name :	<i>C. Pascoe</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513400

Version : 03

Date of issuing : November 28th 2011

CERTIFICATE OF IRRADIATION

Number: **B**

0532570 00

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom

Telephone: +44 (0)1274 686011

Facsimile: +44 (0)1274 686061

Email: assistance@synergyhealthplc.com

Website: www.synergyhealthplc.com

LA1 4XS

P018609

Cust. Ref: 22/05/12

Date Rec'd 24/05/12

Date: 24/05/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B212050761	JD IPA BOTTLES 25-45kGy	709	B/No.W032134 ROUTINE DOSE 37.3, 38.9, 37.1, 36.9 kGy
B2FLE0010999 B212050761	SAMPLES	5	CUSTOMER BATCH/CODE No.s AS ORDER No.P018609 DATED 21/5/12
Despatch Text: 14PLTS			
Total		714	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *J. Coffey*

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 28 May 2012
Date Tested: 31 May 2012
Date Test Completed: 31 May 2012
Purchase Order: 18635

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004950/1
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 28/05/2012
Test Article: Clearklens IPA Block 1
Sample Code: -
Batch Ref: WO32134
Qty Received: 1 x 900mL

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	99 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 06 Jun 2012 12:44:05

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology



Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8
 Hightown
 White Cross Industrial Estate
 Lancaster
 LA1 4XS

Date Received: 28 May 2012
Date Tested: 12 Jun 2012
Date Test Completed: 26 Jun 2012
Purchase Order: 18635

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004949/1
Test Required: Sterility Test by Membrane Filtration (Steritest)
Date Received: 28/05/2012
Test Article: Clearklens IPA Block 1
Sample Code: -
Batch Ref: WO32134
Qty Received: 20 x 900mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	11 x 50mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

See Deviation Report PHM 2012 080, Investigation report PHM2012 090

Approval is provided by Electronic Signature. Their name is shown below.

Date: 28 Jun 2012 13:01:20

Mrs P. Pham-Lengoc, BSc, MSB
 Laboratory Manager - Pharmaceutical Microbiology