



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 18/6/12

Product Name	CLEARLEN S TEGO 2000SC VH25
Product Code	7513387
Batch Number	W032136
Production Date	05/2012
Expiration Date	05/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow	Clear slightly yellow liquid
pH 100% (20°C)	JDM001	7.5	8.5	7.5
Specific Gravity (20°C)	JDM004	0.993	1.003	1.002

On behalf of Diversey site Quality Manager	Name :	<i>C. Piscoe</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com

LA1 4XS
 P018649
 31/05/12
 8/06/12

Cust. Ref: P018649
 Date Rec'd 31/05/12
 Date: 8/06/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B212060186	DIVERSEY TEGO 2000SC 50ml 25-42kGy	105	B/No.W032136 - CUST/BN0.402760 ROUTINE DOSE 34.3, 34.3, kGy
B2FLE0010999 B212060186	SAMPLES	1	CUSTOMER BATCH/CODE No.s AS ORDER No.P018649 DATED 30/5/12
B2FLE001MIM	MINIMUM CHARGE UPLIFT		
	Despatch Text: 2PLTS		
	Total	106	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *J. Coffey*
 For and on behalf of SYNERGY HEALTH PLC



Wickham Laboratories

contract analytical services

Wickham Laboratories Limited,
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
+44(0)1329 832511
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 18 Jun 2012
Date Tested: 06 Jul 2012
Date Test Completed: 28 Jul 2012
Purchase Order: 18696

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0005308/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: PH Eur and USP
Date Received: 18/06/2012
Test Article: TEGO 2000sc
Sample Code: 402760
Batch Ref: Works Order 32136
Qty Received: 20 x 50mL Bottles + 20 spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Comments

Corrected Certificate of Analysis to LR0005308/1 authorised 31 Jul 2012 16:59:54; sample code and batch reference numbers mixed up.

As Steritest Sterility Test

Approval is provided by Electronic Signature. Their name is shown below.

Date: 26 Apr 2013 13:45:09

Mrs S.L. Wood, BSc (Hons), CBIol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0005308

Certificate of Analysis - OSMM

Print Number: P0010133

Page 1 of 1