

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

### **CERTIFICATE OF ANALYSIS**

Date: 29/06/2018

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514862
Batch Number	FMP 18178 45757
<b>Production Date</b>	271061 2018
<b>Expiration Date</b>	EXP 27/06/2020

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Liı	mits	Results
	<u> </u>	Lower	- Upper	
Appearance	Visual		olourless quid	Clear Colour less Liquid
pH (neat solution)	DM001	1.5	2.5	1.5
Specific Gravity (20°C)	DM004	1.010	1.030	1.016

On hele of Diverse with	Name:	Partyrisho, Vandanda
On behalf of Diversey site Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version : 02 Date of issuing : November 24<sup>th</sup> 2017



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

#### **CERTIFICATE OF ANALYSIS**

Date: 29/06/2018

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	FMP 18178 45757
<b>Production Date</b>	27/06/2018
Expiration Date	EXP 27/06/ 2020

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lir	nits	Results
		Lower -	0 0 0 0 0 0	
Appearance	Visual	Clear Colou	rless Liquid	Clear Colovia
pH (neat solution)	DM001	9.0	12.5	11.7
Specific Gravity (20°C)	DM004	1.004	1.020	1.008
Potential Chlorine Dioxide (ppm)		100	180	151.875

O-1-1-15-6Di	Name:	Partyista lavoluda
On behalf of Diversey site  Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

The Botonicit Benig issued electronical	,,	
COA Template	Version: 02	Date of issuing : November 24 <sup>th</sup> 2017



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: Fax:

+44(0)1329 226600 +44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

Date Received:

05 Jul 2018

Date Tested:

10 Jul 2018

**Date Test Completed:** 

24 Jul 2018

Purchase Order:

31329

### **CERTIFICATE OF ANALYSIS**

**Laboratory Reference Number:** 

0043549/1

**Test Required:** 

Sterility by Membrane Filtration Steritest

Date Received:

05/07/2018

Test Article:

**BI-Spore Base** FMP18178 45757

Sample Code: Batch Ref:

610100

**Qty Received:** 

20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 41 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM05

Approval is provided by Electronic Signature. Their name and position is shown below.

1277

CBio MRSB

Date: 25 Jul 2018 10:51:39

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0043549 Print Number: P0052283

Page 1 of 1





Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600

Fax:

+44(0)1329 226688

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA 1 4XS

Date Received:

05 Jul 2018

Date Tested:

10 Jul 2018

**Date Test Completed:** 

24 Jul 2018

Purchase Order:

31329

## **CERTIFICATE OF ANALYSIS**

Laboratory Reference Number:

0043549/2

**Test Required:** 

Sterility by Membrane Filtration Steritest

**Date Received:** 

05/07/2018

Test Article: Sample Code:

**BI-Spore Activator** FMP18178 45757

**Batch Ref:** 

610101

**Qty Received:** 

20 x 100mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 9.0 2.6.1, USP 41 <71> & JP XVII 4.06
Product Standard Data Sheet	FG047/psd	FM04

Approval is provided by Electronic Signature. Their name and position is shown below.

CBio(M858

Date: 25 Jul 2018 10:51:39

Mrs C Moore

Laboratory Manager - Pharmaceutical Microbiology

Certificate of Analysis - OSMM Consignment: 0043549 Print Number: P0052283

Page 1 of 1





http://www.steris-ast.com

# **Certificate of Irradiation**

Date Issued: 21-May-2018

UK33S12104781-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order information			
Account Number:	100432		
Synergy Health Sales Part Reference:	1106825		
Customer Reference Number:	P031016		
Product Description:	BI-SP B3 COMPONENTS DV4725 25-40kGy		
Validation Reference:	4725		
Quantity Received:	16		
Customer Minimum Specification kGy:	25.0		
Customer Maximum Specification kGy:	40.0		
Customer Unit Lot/Batch Number:	W045392, 2 PLTS		
	rradiation Data		
Date and Time of Irradiation:	18-May-2018 08:30		
Reference Dose Range kGy:	31.7 - 33.4		
Calculated Minimum Dose kGy:	26.3		
Calculated Maximum Dose kGy:	34.1		

Items irradiated under WO45392, will be used in finished product batch FMP18178 45757.

C Pascoe

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company