



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 09-01-2019

Product Name	DI Clearklens IPA 12x500mL
Product Code	7519268
Batch Number	MUL 461822
Manufacturing Date	04-01-2019
Expiration Date	04-01-2021

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

The Manufacturing/Expiration dates refer to packing date of the bulk product. As the packaging process could last for more than 1 day, a difference between the dates mentioned on this document and those printed on delivered stock is possible. For stock management purposes it is recommended that the dates printed on the packs are used.

Test	Test Method	Limits	Results
		Lower - Upper	
Appearance / Colour	MF01	Colorless Liquid	OK
Odour	MF02	Alcohol	OK
PH (Neat)	MF03 / DM-001	6,00 - 9,00	9,00
Specific Gravity (g/ml at 20°C)	MF08 / DM-004	0,8699 - 0,8798	0,8750

Manufacturer: Multifill BV Constructieweg 25A 3641 SB Mijdrecht The Netherlands	Name :	J. Koppenol
	Position	Quality Control Inspector

JK

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 th 2017
--------------	--------------	--