

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

27 January 2021

Product Name	CLEARKLENS IPA VH1					
Product Code	101100418					
Batch Number	FMP21004	51120				
Production Date	04/01/2021	V1120				
Expiration Date	EXP 04/01/2023					

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method Limits Lower - Upper				
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liqui	
Specific Gravity (20°C)	DM004	0.872	0.883	0.875	

On behalf of Diversey site	Name:	urszula Haraburda Ewelina Markiel			
Quality Manger	Position	Quality Control Inspector			

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 01/18/2021 11:14:18 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2806536 P.O.# PO37152 Sales Order # 2672869

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 01/15/2021 17:04:14 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	482 BX	101101081	Diversey Clearklens DE 900ml	FMP20357	WO51119/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP20357	WO51119/DEV002096
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	375 BX	101100418	Diversey Clearklens IPA/DE 900	FMP21004	WO51120/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	860 BX	Total			

Quality Test Summary

				or ourminar,	•	Signed By	-
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.5 KGY	Pass	SWARD SIMON WARD	01/16/2021 05:08:03 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.5 KGY	Pass	SWARD SIMON WARD	01/16/2021 05:08:35 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 01/18/2021 11:08:13 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

01/18/2021 11:04:46 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2808794
P.O.# PO37173 Sales Order # 2675418

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 01/16/2021 02:22:54 GMT

SO Line #	Qty l	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	335 E	вх	101100418	Diversey Clearklens IPA/DE 900	FMP21004	WO51120/DEV002080
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	5 E	BX	101100418_SAMPLE	Diversey Clearklens IPA/DE 900	FMP21004	WO51120/DEV002080
	Dose Map		302_0030	Ref Pos: 10E1, AF Min = 0.71		
	340 E	вх	Total			

Quality Test Summary

					,	Signed By	_
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.6 KGY	Pass	DMARRIOTT DAVID MARRIOTT	01/16/2021 15:57:30 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	38.9 KGY	Pass	AHOBSON ASHLEY HOBSON	01/18/2021 10:40:52 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 01/18/2021 11:01:23 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21004 51120

Lucideon Sample Number: UK21349-2340

Lucideon Report Number: UK21349-2340/ETEP

Issue Number:

Date Logged:

27-Jan-2021

Order Number:

PO 37291

Date Reported:

09-Feb-2021

Date(s) of Test(s):

09-Feb-2021 to 09-Feb-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

15-49-PD 100801

Mrs Natalie Boot

Business Support Administrator

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21004 51120

Lucideon Sample Number: UK21349-2339

Lucideon Report Number: UK21349-2339/MFEP

Issue Number:

1

Date Logged:

27-Jan-2021

Order Number:

PO 37291

Date Reported:

22-Feb-2021

Date(s) of Test(s):

06-Feb-2021 to 21-Feb-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

0BOOK 22-Feb-21

Mrs Natalie Boot

Business Support Administrator