

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

13 May 2021

Product Name	CLEARKLENS IPA VH1			
Product Code	7513400			
Batch Number	FMP21091	51421		
Production Date	01/04/2021			
Expiration Date	EXP 01/04/2023			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits Lower - Upper		Results	
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liqu	
Specific Gravity (20°C)	DM004	0.872	0.883	0.875	

On behalf of Diversey site	Name:	Ewelina Markiel Justyna Staron			
Quality Manger	Position	Quality	Control Inspector		

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

05/04/2021 12:20:32 GMT

Page 1 of 1

Cianad Die

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2872301 P.O.# PO37896 Sales Order # 2753269

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 05/01/2021 13:35:04 GMT

so UOM **Customer Item Number Customer Item Description** Customer Qty Customer Line # Lot Number Load Number 101.000 960 ВХ 7513400 Diversey Clearklens IPA 900 mL FMP21091 WO51421/DEV001152 Dose Map 302 0030 Ref Pos: 10E1, AF Min = 0.71 вх 960 Total

Quality Test Summary

						Signea By	•
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.3 KGY	Pass	IWEATHERAL	05/02/2021 05:40:15 GMT
		Reason Code Test				IAN WEATHERALL	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.2 KGY	Pass	IWEATHERAL	05/02/2021 05:40:47 GMT
		Reason Code Test				IAN WEATHERALL	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 05/04/2021 12:20:22 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

05/10/2021 10:11:20 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2873070 P.O.# PO37898 Sales Order # 2754220

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 05/08/2021 03:26:54 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	457 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21091	WO51421/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21091	WO51421/DEV001152
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
103.000	440 BX	100862174	Diversey Clearklens DE 900 mL	FMP21110	WO51686/DEV001910
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	900 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.5 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:48:45 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	39.7 KGY	Pass	JBARRATT JAMIE BARRATT	05/08/2021 16:49:06 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 05/10/2021 10:11:01 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21091 51421

Lucideon Sample Number: UK212095-14648

Lucideon Report Number: UK212095-14648/ETEP

Issue Number:

Date Logged:

13-May-2021

Order Number:

PO 38044

Date Reported:

Date(s) of Test(s):

25-May-2021

Endotoxin Test (EP)

20-May-2021 to 20-May-2021

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

End of Test Report



Mrs Andrea Saunders

Business Support Administrator

Lucideon Limited

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21091 51421

Lucideon Sample Number: UK212095-14647

Lucideon Report Number: UK212095-14647/MFEP

Issue Number:

Date Logged:

13-May-2021

Order Number:

PO 38044

Date Reported:

Date(s) of Test(s):

11-Jun-2021

20-May-2021 to 03-Jun-2021

Sterility Testing Membrane Filtration EP

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

End of Test Report

CBOOK 11-JUN-21

Mrs Natalie Boot

Business Support Administrator