



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 23/03/2022

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	100848254
Batch Number	FMP22082 54330
Production Date	23/03/2022
Expiration Date	EXP 23/09/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid	11.4	Clear Slightly Yellow Liquid
pH (neat solution)	DM001	9.0	11.4	10.8
Specific Gravity (20°C)	DM004	0.990	1.010	0.997

On behalf of Diversey site Quality Manager	Name :	Justyna Staron, Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of Issuing : November 24 th 2017
--------------	--------------	--

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-26633A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
C/SINALD SS 900ml Bottle DV4673	54469	13	Case
Validation Reference Number: 4673			

Processing Run Start Date: 04-Mar-2022 6:47 PM

Processing Run End Date: 05-Mar-2022 1:52 AM

Specified Dose Range (kGy):	25.0 - 40.0	Calculated Min Dose (kGy):	29.6
Reference Dose Range (kGy):	27.3 - 36.6	Calculated Max Dose (kGy):	36.1

PO Number: 402084

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 05-Mar-2022 8:38 AM

□

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

Items irradiated under WO54469 will be used in finished batch FMP22082 54330.

Olga Kirchner

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Cleansinald SS

Sample Description: Sample Code: FMP22082 54330

Lucideon Sample Number: UK221489-9429

Lucideon Report Number: UK221489-9429/MFEP **Issue Number:** 1

Date Logged: 07-Apr-2022 **Order Number:** 402354

Date Reported: 03-May-2022 **Date(s) of Test(s):** 13-Apr-2022 to 27-Apr-2022

Sterility Testing
Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

NcBoor 03-May-22

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com