



Diversey Europe BV

Certificate of Analysis

Date 18.05.2022
Purchase order item/date 4500251906 / 25.02.2022
Delivery item 87325740 900001
Order item/date 3913903 000010 / 25.02.2022
Customer number 5215646

Company
Henri Essers en Zonen Internationaa
Transport NV
Henry Fordlaan 8
3600 GENK
BELGIUM

Repeat printout

Page
1 / 1

Article
70001369 ClearKlens Oxifast -sterile- 1 l FL

Batch: SEP/1581341/22/060
Expiry Date: 2024/03
Date of manufacture: 01/03/2022

Characteristic Method	Unit	Value	Lower Limit	Upper Limit	Conformity
peracetic acid MAN LCQ 011	%	0,09	0,07	0,10	Yes
Hydrogen peroxide MAN LCQ 011	%	3,04	2,90	3,40	Yes
loss OAT 16h at 96 ° C MAN LCQ 009	%	0,45	0,00	2,50	Yes
stérility certificate european pharmacopea 2.6.1			conform		Yes

released on: 17/05/2022 by: PUGEAUD

Comments:

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the MSDS.

This certificate of analysis has been printed through an EDP-System and is therefore unsigned.

***End of certificate of analysis ***