

## CERTIFICATE OF ANALYSIS

Date issue: 23/01/2023


<b>Product Name</b>	DI Tego 2000 S 50x0.05L
<b>Product Code</b>	101107171
<b>Batch Number</b>	FMP23003 55625
<b>Production date</b>	03/01/2023
<b>Expiration Date</b>	EXP 03/01/2026

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Conformity
Appearance	Visual	Colorless Liquid Clear	N/A	Colorless Liquid Clear
Specific Gravity (20°C)	DM 004	0.993 – 1.003	g/cm <sup>3</sup>	1.000
pH (neat Solution)	DM 001	7.5 – 8.5	N/A	8.3

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 45.0 (kGy)	YES	2173-34127A
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	Pass	UK23272-1906/MFEP

	Date of released On behalf of Diversey's Manufacturing location: 09/03/2023	by :  Position: QA Assistance
---	--	-------------------------------------

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

\*\*\*End of certificate of analysis\*\*\*

# STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-34127A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
TEGO 2000SC DV4724	FMP23003 - 55625	93	Case
Validation Reference Number: 4724			
TEGO 2000SC DV4724	FMP23003 - 55625 1 SAMPLE BOX	1	Case
Validation Reference Number: 4724			

Processing Run Start Date: 17-Jan-2023 5:42 AM

Processing Run End Date: 17-Jan-2023 3:38 PM

Specified Dose Range (kGy):	25.0 - 45.0	Calculated Min Dose (kGy):	28.7
Reference Dose Range (kGy):	29.4 - 40.5	Calculated Max Dose (kGy):	37.7

PO Number: 404283

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

**Gamma Process Run Approval authorized by STERIS**

Date/Time E-Signed: 17-Jan-2023 8:33 PM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

## Processing Location

Synergy Health Sterilisation UK  
Limited, a STERIS Company  
Brunel Close  
Drayton Fields Industrial Estate  
Daventry  
Northants  
NN11 8RB  
Phone: + 44(0) 1327 706 111

# PHARMACEUTICAL ANALYSIS REPORT

**LUCIDEON**

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Tego 2000SC

**Sample Description:** Sample Code: Fmp23003 55625

**Lucideon Sample Number:** UK23272-1906

**Lucideon Report Number:** UK23272-1906/MFEP

**Issue Number:** 1

**Date Logged:** 24-Jan-2023

**Order Number:** 404376

**Date Reported:** 14-Feb-2023

**Date(s) of Test(s):** 26-Jan-2023 to 09-Feb-2023

### Sterility Testing

Membrane Filtration EP

## Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report**

*Natalie Boot 14 Feb 23*

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com