

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 16 February 2023

Product Name	CLEARKLENS B	CLEARKLENS BI-SPORE BASE SOLUTION VH26 first off		
Product Code	7514852			
Batch Number	FMP23047	55738		
Production Date	16/02/2023			
Expiration Date	EXP 16/02/2025			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	- Upper	
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	1.9
Specific Gravity (20°C)	DM004	1.010	1.030	1.018

On behalf of Diversey site	Name:	Justyna Staron	Edyta Rodrígues
Quality Manger	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA Template Version: 04 Date of issuing: March 11th	2019	
--	------	--



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 20 February 2023

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
Product Code	7514852		
Batch Number	FMP23047 55738		
Production Date	16/02/2023		
Expiration Date	EXP 16/02/2025		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	11.5
Specific Gravity (20°C)	DM004	1.004	1.020	1.011
Potential Chlorine Dioxide		100	180	145.125

On behalf of Diversey site	Name:	Justyna Staron	Edyta Rodrígues
Quality Manger	Position	Quality Control Inspector	

This document being issued electronically does not bear a signature

COA Template Version: 04 Date of issuing: March 11th 2019

STERIS: Gamma Certificate Of Processing

Prepared for:

FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID

2173-33390A

Product Code

Lot Number

Quantity

BI-SP B3 Components DV5857

55368

UOM

Case

Validation Reference Number:

5857

Processing Run Start Date: 10-Dec-2022 1:24 PM

Processing Run End Date: 10-Dec-2022 9:04 PM

Calculated Min Dose (kGy):

29.9

Specified Dose Range (kGy): Reference Dose Range (kGy): 25.0 - 40.0 26.1 - 35.8

Calculated Max Dose (kGy):

35.4

PO Number: 404108

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 11-Dec-2022 7:13 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company **Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB**

Phone: + 44(0) 1327 706 111

Items irradiated under WO55368 will be used in finished batch Fmp23047 55738.

Olga Kirchner

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp23047 55738

Lucideon Sample Number: UK23698-4719

Lucideon Report Number: UK23698-4719/MFEP

Issue Number:

Date Logged:

24-Feb-2023

Order Number:

404638

02-Mar-2023 to 16-Mar-2023

Date Reported:

17-Mar-2023

Date(s) of Test(s):

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

MBOOK ITMORES

Mrs Natalie Boot

Senior Business Support Administrator

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp23047 55738

Lucideon Sample Number: UK23698-4720

Lucideon Report Number: UK23698-4720/MFEP

Issue Number:

Date Logged:

24-Feb-2023

Order Number:

404638

02-Mar-2023 to 16-Mar-2023

Date Reported:

17-Mar-2023

Date(s) of Test(s):

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



Mrs Natalie Boot

Senior Business Support Administrator