


SIMA PHARMA	Form	SP-PR-LIB-F-014	Status : APPLICABLE
	Certificate of Analysis DIVERSEY for ClearKlens Cleansinald RTU VH9S 10x1L		Date of application : 13/07/2023
			Index : 01

	ClearKlens Cleansinald RTU VH9S 10x1L SKU 100848252		
Batch Number	SPH20072023		
Production date	07/2023		
Expiration Date	12/2024		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

ClearKlens Cleansinald RTU VH9S				
Test	Test Method	Specification	Unit	Results
Appearance	PS-CON-I-029	Clear colorless liquid	N/A	clear colorless liquid
Relative density (20°C)	PS-CON-I-027	0.990 - 1.010	-	0.999
pH (neat Solution)	PS-CON-I-040	9.0 - 11.4	-	10.3
Content of cationic substances (%)	PS-CON-I-034	0.070 - 0.090	%	0.076


Water routine test				
Microbiological Quality of WFI Water	Water PPI VRAC European pharmacopoeia	<10 cfu/100ml	CFU / 100mL	<1

Control of the Sterility of the product:


Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 - 45.0 (kGy)	C	2013-1120A + FR02S12746963 -1-1
Sterility certificate	Membrane filtration sterility testing EP 11.0 2.6.1	No Growth	C	23/CBH/STE/208

The analysis results above could change over time in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the Safety Data Sheet.

SIMA PHARMA	Form	SP-PR-LIB-F-014	Status : APPLICABLE
	Certificate of Analysis DIVERSEY for ClearKlens Cleansinald RTU VH9S 10x1L		Date of application : 13/07/2023
			Index : 01

	ClearKlens Cleansinald RTU VH9S 10x1L SKU 100848252		
Batch Number	SPH20072023		
Production date	07/2023		
Expiration Date	12/2024		

We certify that, except for the exceptions or deviations listed above, the quoted supply has been manufactured and tested in accordance with the requirements of the specifications in force

SIMA-PHARMA Control (Responsible for the production process of the final product) :	Stamp and signature
Name: E. JACOIN Position: General Manager Produced on behalf of Diversey at : Sima Pharma: Z.I. de Rousset / Peynier, 54 Av. de la Plaine, 13790 Rousset, France Date : 11/09/23	

End of certificate of analysis

For further information, please contact Diversey at pharma@diversey.com