Diversey

Diversey Europe BV

Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 11/12/14

Product Name	CLEARKLENS IPA VH1 7514372	
Product Code		
Batch Number	FMP14321 37088	
Production Date	11/2014	
Expiration Date	11/2016	

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower -		
Appearance	Visual	Clear C	olourless	Clear Colouriess
		Lie	quid	1,20d
Specific Gravity (20°C)	JDM004	0.872	0.883	0.875

On behalf of Diversey site Quality Manager	Name :	CASOC, OGO kinchna
	Position	Quality Control Inspector

This document being issued electronicall	y does not bear a signature	
COA 7514372	Version: 03	Date of issuing: November 28th 2011



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 04-Dec-2014

UK32S11286315-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd Unit 8, White Cross Ind Estate Hightown Lancaster Lancashire LA1 4XS UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information **Account Number:** 100432 Synergy Health Sales Part Reference: 1002734 **Customer Reference Number:** P023368 Product Description: JD IPA BOTTLES 25-45kGy Validation Reference: 4.1723 **Quantity Received:** 708 Customer Minimum Specification kGy: 25.0 Customer Maximum Specification kGy: 45.0 Customer Unit Lot/Batch Number: 704 X FMP 14321 37088 + 4 SAMPLES THURS 27.11.14 13PLTS LOT 3 irradiation Data Date and Time of Irradiation: 03-Dec-2014 20:12 Reference Dose Range kGy: 36.9 - 38.7 Calculated Minimum Dose kGy: 26.0 Calculated Maximum Dose kGy: 41.4



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone: +44(0)1329 226600 Alt. Telephone: +44(0)1329 832511 Fax: +44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

08 Dec 2014

Date Tested:

16 Dec 2014

Date Test Completed:

30 Dec 2014

Purchase Order:

23416

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0020179/1

Test Required:

Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Date Received:

08/12/2014

Test Article:

IPA ClearKlens Block 2

Sample Code:

6024

Batch Ref:

FMP14321 37088

Qty Received:

20 x 900mL Bottles + 1 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20- 25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested Product Standard Data Sheet	MM107/05 FG047/psd	20 x 50 mL FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

amore

Date: 31 Dec 2014 14:28:21

Mrs C Moore

ISO 17025

Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0020179

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0024744

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire



Wickham Laboratories Ltd Hoeford Point, Barwell Lane, Gosport Hampshire PO13 0AU England

Telephone:

+44(0)1329 226600

Alt. Telephone:

+44(0)1329 832511

Fax:

+44(0)1329 832511

mail@wickhamlabs.co.uk www.wickhamlabs.co.uk

Ms D Henderson

Flexible Medical Packaging Limited

Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA1 4XS Date Received:

08 Dec 2014

Date Tested:

10 Dec 2014

Date Test Completed:

10 Dec 2014

Purchase Order:

23416

CERTIFICATE OF ANALYSIS

Laboratory Reference Number:

0020179/2

Test Required:

Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

08/12/2014

Date Received: Test Article:

IPA ClearKlens Block 2

Sample Code:

6024

Batch Ref: Qty Received:

FMP14321 37088 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.2 EU/mL
Spiked Recovery	MM110/01	112 %
Dilution Tested	MM110/04	1/40
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM02

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Eur Wood.

Date: 10 Dec 2014 16:03:51

Mrs S.L. Wood

Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0020179

GLP GMP FDA

Certificate of Analysis - OSMM

Print Number: P0024356

Page 1 of 1

Company Registered in England No. 752951 Wickham Laboratories Limited, Registered Office: Hoeford Point, Barwell Lane, Gosport, Hampshire