



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 22-08-2015

Product Name	CLEARKLENS DE VH29
Product Code	7515783
Batch Number	FMP 15197 38577
Production Date	16/07/2015
Expiration Date	16/07/2017

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		clear Colourless Liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.883

On behalf of Diversey site Quality Manager	Name :	Pantziola, Alina
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7515783	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 09-Aug-2015

UK32S11435608-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
Unit 8, White Cross Ind Estate
Hightown
Lancaster
Lancashire LA1 4XS
UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002733
Customer Reference Number:	P024822
Product Description:	JD DE 900ML BOTTLES 25-45kGy
Validation Reference:	4.1756
Quantity Received:	701
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	697 X B/N FMP15197 38577 + 4 SAMPLES TUES 28TH JULY 2015 13 PLTS LOT 1

Irradiation Data

Date and Time of Irradiation:	08-Aug-2015 22:39
Reference Dose Range kGy:	37.7 - 40.5
Calculated Minimum Dose kGy:	28.3
Calculated Maximum Dose kGy:	43.9

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 18 Aug 2015
Date Tested: 20 Aug 2015
Date Test Completed: 20 Aug 2015
Purchase Order: PO24935

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024273/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 18/08/2015
Test Article: ClearKlens DE VH29
Sample Code: 6773
Batch Ref: FMP15197 38577
Qty Received: 1 x 900mL Bottle

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	97 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 21 Aug 2015 16:32:24

Mr K.A. Barker
Business Manager - Microbiology



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 18 Aug 2015
Date Tested: 21 Aug 2015
Date Test Completed: 04 Sep 2015
Purchase Order: PO24935

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0024273/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 18/08/2015
Test Article: ClearKlens DE
Sample Code:
Batch Ref: FMP15197 38577
Qty Received: 20 x 900mL Bottles

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

The testing procedure was performed in compliance with the current Pharmacopoeias however, the volume tested does not comply with the requirements of the current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 07 Sep 2015 15:33:36

Mrs C Moore
Laboratory Manager - Pharmaceutical Microbiology