

**CERTIFICATE OF IRRADIATION** Number: **B** 0381280 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
Unit B  
White Cross Industrial Estate  
Hightown  
Lancaster



**Isotron Limited**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

LA1 4XS

Cust. Ref: PO13083  
Date Rec'd 28/08/08  
Date: 2/09/08

| ITEM CODE<br>ISOTRON BATCH | ITEM SPECIFICATION                  | QTY | ADDITIONAL DETAILS  |
|----------------------------|-------------------------------------|-----|---|
| B2FLE0010091<br>B208080748 | DIVERSEY TEGO2001 10X5x100 25-38kGy | 188 | B/No.WO24716 CUSTOMER B/No.MK266405<br><br>ACTUAL....33.9 - 35.0....kGy<br>DOSE |
| B2FLE0010999<br>B208080748 | SAMPLES                             | 1   | CUSTOMERS BATCH/CODE No.s AS ORDER<br>No.PO13083 DATED 27/8/08                  |
| Total                      |                                     | 189 | Last Page 1 of 1  |

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

Date Received : 10-Sep-08  
Date Tested : 16-Sep-08  
Date Completed : 30-Sep-08  
Purchase Order No. : P013130

Test Article : Clearklens Tego 2001SC  
Sample Code : Works Order: 24716  
Batch Reference : MK00266405  
Laboratory Reference : LR-00788058 Article Quantity : 20 x 100mL  
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

| Test   |   | Method   | Result |
|--|---|----------|--------|
| Sterility Test by Membrane Filtration (Steritest) Method | * | MM107    | Pass   |
| Growth in Tryptone Soya Broth at 22°C after 14 days      | * | MM107/01 | 0/2    |
| Growth in Fluid Thioglycollate at 32°C after 14 days     | * | MM107/02 | 0/2    |
| Product Standard Data Sheet                              | * | MQ002/01 | FM03   |

Day Book Ref: 226044  
Test carried out according to current Pharmacopoeias.



*W.A. Barker*

Date 3 Oct 2008

K.A. Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

JohnsonDiversey  
Clean is just the beginning



CERTIFICATE OF ANALYSIS

|                 |                              |
|-----------------|------------------------------|
| PRODUCT NAME    | CLEARKLENS TEGO 2001 SC VH24 |
| PRODUCT CODE    | 7513402                      |
| BATCH NUMBER    | W024716                      |
| PRODUCTION DATE | 08 / 2008                    |
| EXPIRY DATE     | 08 / 2010                    |

| TEST                    | METHOD     | LIMITS                            | RESULTS          |
|-------------------------|------------|-----------------------------------|------------------|
| Appearance              |            | Clear Colourless to Yellow Liquid | Yellowish Liquid |
| Odour                   |            | Typical                           | Typical          |
| pH (neat)               | GM 0130 01 | 8.5 - 9.5                         | 8.6              |
| Density (20°C)          | GM 0110 01 | 1.060 - 1.070                     | 1.062            |
| Refractive Index (20°C) | GM 0120 01 | 1.374 - 1.384                     | 1.381            |

Date: 08/09/08

QA Manager: [Signature]

QA: [Signature]

JD Approval: [Signature] 8/9/08

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

Issue no. 5  
Issued by W. Ashton  
Issue date 31.03.2006