


**CERTIFICATE OF ANALYSIS**

<b>PRODUCT NAME</b>	<b>CLEARKLENS CLEANSINALD VH09 SC</b>
<b>PRODUCT CODE</b>	7516429
<b>BATCH NUMBER</b>	W025159
<b>PRODUCTION DATE</b>	10-2008
<b>EXPIRY DATE</b>	10-2010

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)		Clear Yellow Solution	clear yellow solution
pH (1% soln)	GM 0130 01	12.0 – 13.0	12.8
Density (20°C)	GM 0110 01	1.04 – 1.06	1.058

Date: 05/11/08

QA Manager: 

QA: L. Gargy -

JD Approval: 

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 06-Nov-08  
**Date Tested** : 11-Nov-08  
**Date Completed** : 25-Nov-08  
**Purchase Order No.** : PO13348


**Test Article** : Cleansinald  
**Sample Code** : Works Order No. 25159  
**Batch Reference** : VLF99C01 08 056 W025159 2010/10  
**Laboratory Reference** : LR-00803664  
**Article Quantity** : 20 Bottles  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM06

Day Book Ref: 306022

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.



*WKB*

Date 4 Dec 2008

K.A.Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

**Sample Booking Number:** CN00128376

Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*

# CERTIFICATE OF IRRADIATION

Number: **B**

0386593 00

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013279  
 Date Rec'd 16/10/08  
 Date: 22/10/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010113 B208100475	JD-BOTTLES/CAPS/BAGS 25-40kGy	4	B/No.W025154 (W025159)  CUSTOMERS BATCH/CODE No.s AS ORDER No.P013279 DATED 15/10/08
Total		4	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED