



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS IPA VH01
PRODUCT CODE	7513400
BATCH NUMBER	W025335
PRODUCTION DATE	02/2009
EXPIRY DATE	02/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Density (20°C)	GM 0110 01	0.865 - 0.885	0.875

Date: 17/02/09

QA Manager: [Signature]

QA: E. Gough

JD Approval: [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel and is valid unsigned.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013692
 Date Rec'd 5/02/09
 Date: 9/02/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010081 B209020197	JD IPA BOTTLES 25-45kGy	175	B/No.W025335 ACTUAL....39.1 - 40.2....kGy DOSE
B2FLE0010999 B209020197	SAMPLES	2	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P013692 DATED 5/2/09
Total		177	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drake*

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS


Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 18-Feb-09
Date Tested : 24-Feb-09
Date Completed : 10-Mar-09
Purchase Order No. : PO 13739

Test Article : Clearklens IPA Block 1
Sample Code :
Batch Reference : W025335
Laboratory Reference : LR-00832064 **Article Quantity** : 23 x 900mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 380029
Volume Tested: 20 x 50mL
Test carried out according to current Pharmacopoeias.



W.A.B.

Date 16 Mar 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00132317

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

CERTIFICATE OF ANALYSIS



 Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8, Hightown
 White Cross Industrial Estate
 Lancaster LA1 4XS

Date Received : 18-Feb-09
Date Tested : 04-Mar-09
Date Completed : 04-Mar-09
Purchase Order No. : PO 13739

Test Article : Clearklens IPA Block 1
Sample Code :
Batch Reference : W025335
Laboratory Reference : LR-00832065 **Article Quantity** : 1 x 900mL
Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2000 EU/mL
Dilution	*	MM110/01	1/40
Spiked Recoveries	*	MM110/02	98% Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM02

This sample was tested at a 1/40 dilution and spiked with endotoxin to check for interference.
 Test carried out according to current Pharmacopoeias.

 Date 12 Mar 2009

 K.A. Barker, C.Biol, M.I. Biol.,
 Business Manager - Microbiology

Sample Booking Number: CN\00132317
 Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited