

CERTIFICATE OF ANALYSIS


PRODUCT NAME	CLEARKLENS CLEANSINALD VH09 SC
PRODUCT CODE	7516429
BATCH NUMBER	W025584
PRODUCTION DATE	06/2009
EXPIRY DATE	06/2011

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)	Visual	Clear Yellow Solution	clear yellow solution
pH (1% soln)	GM 0130 01	12.0 - 13.0	12.2
Density (20°C)	GM 0110 01	1.04 - 1.06	1.056

Date:.....10/06/09.....

QA Manager:..........

QA:.....E. Galagan.....

JD Approval:..........

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014058
 Date Rec'd 14/05/09
 Date: 20/05/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
BZFLE0010113 B209050431	JD-BOTTLES/CAPS/BAGS 25-40kGy	2	B/No.W025583 - W025584 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014058 DATED 13/5/09
BZFLE001MIN	MINIMUM CHARGE UPLIFT		
Total		2	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *R. Layden*

For and on behalf of ISOTRON LIMITED

CERTIFICATE OF ANALYSIS


Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8, Hightown
 White Cross Industrial Estate
 Lancaster LA1 4XS

Date Received : 03-Jun-09
Date Tested : 15-Jun-09
Date Completed : 29-Jun-09
Purchase Order No. : 014126

Test Article : Clearklens Cleansinald SC
Sample Code :
Batch Reference : WO25584 2011/06
Laboratory Reference : LR-00861254 **Article Quantity** : 20 bottlesx25mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM06

Day Book Ref: 574000
 Volume Tested: All
 Test carried out according to current Pharmacopoeias.



Date 02 July 2009

K.A. Barker, C.Biol, M.I. Biol.,
 Business Manager - Microbiology

Sample Booking Number: CN\00136450

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited