



**CERTIFICATE OF ANALYSIS**

<b>PRODUCT NAME</b>	<b>CLEARKLENS TEGO 2001 SS VH24</b>
<b>PRODUCT CODE</b>	<b>7513403</b>
<b>BATCH NUMBER</b>	W025860
<b>PRODUCTION DATE</b>	03/2009
<b>EXPIRY DATE</b>	01/2010

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless to Pale Yellow Liquid	clear colourless liquid
Odour		Typical Slight	Typical
pH (neat)	GM 0130 01	8.5 - 9.2	8.7
Density (20°C)	GM 0110 01	0.998 - 1.001	1.007

Date:.....27/03/09.....

QA Manager:.....*[Signature]*.....

QA:.....*E. Gaujeu*.....

JD Approval:.....*[Signature]*.....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P013843  
 Date Rec'd 19/03/09  
 Date: 23/03/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B209030565	TEGO 2001 SS 25-42kGy	94	B/No.W025860 - CUSTOMER B/No.283996  ACTUAL....38.8 - 39.7....kGy DOSE
B2FLE0010999 B209030565	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P013843 DATED 18/3/09
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
Total		96	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS


Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 30-Mar-09  
**Date Tested** : 02-Apr-09  
**Date Completed** : 16-Apr-09  
**Purchase Order No.** : 13881

**Test Article** : Clearklens Tego SS Block 1  
**Sample Code** : Works Order: 25860  
**Batch Reference** : 283996  
**Laboratory Reference** : LR-00842474 **Article Quantity** : 20Bottlesx900mL  
**Test Required** : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref:484004  
Volume Tested: 20x50mL  
Test carried out according to current Pharmacopoeias.

  
Mrs C.L. Walker BSc, C.Biol., MIBiol  
Laboratory Manager - Pharmaceutical Microbiology

Date 27 April 2009



K.A.Barker, C.Biol, M.I. Biol.,  
Business Manager - Microbiology

**Sample Booking Number:** CN\00133768  
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*