



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W026130
PRODUCTION DATE	07/2009
EXPIRY DATE	07/2011

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	Yellow clear liquid
pH (neat)	GM 0130 01	7.5 – 8.5	7.6
Density (20°C)	GM 0110 01	0.993 – 1.003	0.999

Date:..... 31/07/09

QA.Manager:..... [Signature]

QA:..... [Signature]

JD Approval:..... [Signature]

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014333
 Date Rec'd 23/07/09
 Date: 27/07/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B209070612	DIVERSEY TEGO 2000 10x5x50 25-38kGy	104	B/No.W026130 ACTUAL....35.7 - 36.4....kGy DOSE CUSTOMERS BATCH/CODE No.s AS ORDER No.P014333 DATED 23/7/09
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
Total		104	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED