


CERTIFICATE OF ANALYSIS


PRODUCT NAME	CLEARKLENS CLEANSINALD VH09 SC
PRODUCT CODE	7516429
BATCH NUMBER	W026228
PRODUCTION DATE	07/2009
EXPIRY DATE	07/2011

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)	Visual	Clear Yellow Solution	clear yellow solution.
pH (1% soln)	GM 0130 01	12.0 - 13.0	12.4
Density (20°C)	GM 0110 01	1.04 - 1.06	1.054

Date:.....21.07.09.....

QA Manager: 

QA:.....

JD Approval:.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS


Date Received : 10-Jul-09
Date Tested : 15-Jul-09
Date Completed : 29-Jul-09
Purchase Order No. : 14269

Test Article : Cleansinald SC Block 1
Sample Code : Works Order No. 226228
Batch Reference : VLF 99 C01 09 140
Laboratory Reference : LR-00871245
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Article Quantity : 50 x 25mL

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM06

Day Book Ref: 574022
Volume Tested: 20x80mL
Test carried out according to current Pharmacopoeias.



WB

Date 03 Aug 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00138027
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

CERTIFICATE OF IRRADIATION Number: **B** 0409762 00

FLE001

*** Reprint ***

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS
Cust. Ref: P014119
Date Rec'd: 28/05/09
Date: 23/03/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010134 B209060051	C/SINALD BUK BOTT/CAPS/BAGS 25-40kGy	10	B/No.M026226 / <i>W026228</i> ^{EG} CUSTOMERS BATCH/CODE No.s AS ORDER No.P014119 DATED 27/5/09 Actual...34.8 - 35.2...kGy Dose Max...35.9...Min...28.9...kGy
B2FLE001VAL	VALIDATION		TRIPLICATE VALIDATION Ref: 4.0616
Total		10	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED