



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS CLEANSINALD VH09 SC
PRODUCT CODE	7516429
BATCH NUMBER	W026234
PRODUCTION DATE	07/2009
EXPIRY DATE	07/2011

TEST	METHOD	LIMITS	RESULTS
Appearance (20°C)	Visual	Clear Yellow Solution	clear yellow solution
pH (1% soln)	GM 0130 01	12.0 – 13.0	12.2
Density (20°C)	GM 0110 01	1.04 – 1.06	1.056

Date: 30/07/09.....

QA Manager: [Signature].....

QA: [Signature].....

JD Approval: [Signature].....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 31-Jul-09
Date Tested : 04-Aug-09
Date Completed : 18-Aug-09
Purchase Order No. : 14374

Test Article : Clearklens Cleansinald Block 1
Sample Code :
Batch Reference : WO26234
Laboratory Reference : LR-00877085 **Article Quantity** : 50 x 25mL
Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM06

Day Book Ref: 574033
Volume Tested: 40 x 25mL
Test carried out according to current Pharmacopoeias.



MR

Date 26 Aug 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN\00138954
Flexible Medical Packaging Limited

Certificate of Analysis - OSMM

Page 1 of 1

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P014266
 Date Rec'd 9/07/09
 Date: 16/07/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010134 B209070385	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy	9	B/No.W026231 - W026234.
B2FLE0010134 B209070385	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy	1	B/No.W026231 I/S.....33.0.....0/S.....37.3....kGy CUSTOMERS BATCH/CODE No.s AS ORDER No.P014266 DATED 8/7/09
Total		10	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED