



CERTIFICATE OF ANALYSIS

PRODUCT NAME	CLEARKLENS DE
PRODUCT CODE	6069643
BATCH NUMBER	100 26599
PRODUCTION DATE	2009 /09
EXPIRY DATE	2011 /09

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Clear Colourless Liquid	Clear Colourless Liquid
Density (20°C)	GM 0110 01	0.876 – 0.896	0.887

Date:.....12/10/09.....

QA Manager:.....*[Signature]*.....

QA:.....*E. Gargan*.....

JD Approval:.....*[Signature]*.....

This document has been produced on behalf of JohnsonDiversey by authorised personnel.

CERTIFICATE OF IRRADIATIONNumber: **B**

0422252 00

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster

**Isotron Limited**

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

Cust. Ref: P014602
Date Rec'd 24/09/09
Date: 2/10/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B209090589	JD DE 900ML BOTTLES 25-45kGy	1415	B/No.W026599 - CUSTOMER B/No.1249 ACTUAL....35.7....37.7....kGy DOSE
B2FLE0010999 B209090589	SAMPLES	2	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P014602 DATED 24/9/09
Total		1417	Last Page 1 of 1

All in accordance with current
technical agreement

This is to certify that the above items have been irradiated as
specified above

Authorising Signature:

A handwritten signature in blue ink, appearing to read "J. Goffey", written over a horizontal line.

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 08-Oct-09
Date Tested : 22-Oct-09
Date Completed : 05-Nov-09
Purchase Order No. : PO14652

Test Article : Clearklens DE Block 1

Sample Code :

Batch Reference : WO26599

Laboratory Reference : LR-00895013

Article Quantity : 20 x 900mL

Test Required : Sterility Method I: Membrane Filtration by Steritest. BP, Ph Eur and USP.

Test		Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	*	MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	*	MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	*	MM107/02	0/1
Product Standard Data Sheet	*	MQ002/01	FM03

Day Book Ref: 649026

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

Sue Wood

Mrs S L Wood, C.Biol., MSB
Laboratory Manager - Pharmaceutical Microbiology

Date 12 Nov 2009

GLP

Compliant
Laboratory

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

Sample Booking Number: CN00141620

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8, Hightown
White Cross Industrial Estate
Lancaster LA1 4XS

Date Received : 08-Oct-09
Date Tested : 12-Oct-09
Date Completed : 12-Oct-09
Purchase Order No. : PO14652

Test Article : Clearklens DE

Sample Code :

Batch Reference : WO26599

Laboratory Reference : LR-00895012

Article Quantity : 1 x 900mL

Test Required : LAL Test - Kinetic QCL. Inhibition/Enhancement Assay. BP, Ph Eur and USP.

Test		Method	Result
Bacterial Endotoxin by LAL, Kinetic QCL Test	*	MM110	<0.2500 EU/mL
Dilution	*	MM110/01	1/50
Spiked Recoveries	*	MM110/02	Valid
Endotoxin Spike Level	*	MM110/09	0.5 EU/mL
Product Standard Data Sheet	*	MQ002/01	FM01

This sample was tested at a 1/50 dilution and spiked with endotoxin to check for interference.
Test carried out according to current Pharmacopoeias.

MB

Date

20 Oct 2009

K.A.Barker, C.Biol, M.I. Biol.,
Business Manager - Microbiology

GLP

Compliant
Laboratory

Sample Booking Number: CN00141620

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited