

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 Cust. Ref: P015103  
 Date Rec'd 11/02/10  
 Date: 16/02/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210020270	DIVERSEY TEGO 2000SC 50ml 25-42kGy	95	76XB/No.W027603 CUSTOMER B/No.319883 19XB/No.W027544 CUSTOMER B/No.303485  ACTUAL DOSE....34.2 - 34.9....kGy
B2FLE0010999 B210020270	SAMPLES	4	2 X B/No.W027603 2 X B/No.W027544  CUSTOMERS BATCH/CODE No.s AS ORDER No.P015103 DATED 10/2/10
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
	Despatch Text: 3PLTS		
	Total	99	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



**CERTIFICATE OF ANALYSIS**

PRODUCT NAME	CLEARKLENS TEGO 2000 SC VH25
PRODUCT CODE	7513387
BATCH NUMBER	W027603
PRODUCTION DATE	02/2010
EXPIRY DATE	02/2012

TEST	METHOD	LIMITS	RESULTS
Appearance	Visual	Yellow Clear Liquid	yellow clear liquid
pH (neat)	GM 0130 01	7.5 - 8.5	7.5
Density (20°C)	GM 0110 01	0.993 - 1.003	1.002

Date:..... 12/03/2010 .....

QA Manager:..... *[Signature]* .....

QA:..... *E. Gaurin* .....

JD Approval:..... *[Signature]* .....

*This document has been produced on behalf of JohnsonDiversey by authorised personnel.*