



Contact

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 7/9/10

<b>Product Name</b>	CLEARKLENS TEGO 2001SC VH24
<b>Product Code</b>	7513402
<b>Batch Number</b>	<u>W028498</u>
<b>Production Date</b>	<u>08/2010</u>
<b>Expiry Date</b>	<u>08/2012</u>

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	<u>clear slightly yellow liquid</u>
pH (100% 20°C)	JDM001	8.5 - 9.5	<u>8.7</u>
Specific Gravity (20°C)	JDM004	1.060 - 1.070	<u>1.062</u>

"SIGNATURE": *C. Pascoe*

Name Caroline Pascoe

Function Name Pharma Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015821  
 Date Rec'd 26/08/10  
 Date: 31/08/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B210080795	DIVERSEY TEG02001 10x5x100 25-42kGy	99	W02849B - CUST B/NO 7515139  ACTUAL DOSE....33.8 - 33.9....kGy
B2FLE0010999 B210080795	SAMPLES	6	CUSTOMERS BATCH/CODE NO'S AS ORDER NO PO 15821 DATED 25.08.10
Despatch Text:  2PLTS			
Total		105	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffen*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8, Hightown  
White Cross Industrial Estate  
Lancaster LA1 4XS

**Date Received** : 17-Sep-10  
**Date Tested** : 30-Sep-10  
**Date Completed** : 14-Oct-10  
**Purchase Order No.** : 15900

**Test Article** : Clearklens Tego 2001 SC VH24

**Sample Code** : 7515139

**Batch Reference** : WO28498

**Laboratory Reference** : LR-00984260

**Article Quantity** : 20x100mL Bottle

**Test Required** : Sterility Method I: Membrane Filtration by Steritest. Ph Eur and USP.

Test	Method	Result
Sterility Test by Membrane Filtration (Steritest) Method	* MM107	Pass
Growth in Tryptone Soya Broth at 22°C after 14 days	* MM107/01	0/1
Growth in Fluid Thioglycollate at 32°C after 14 days	* MM107/02	0/1
Product Standard Data Sheet	* MQ002/01	FM03

Day Book Ref: 000019

Volume Tested: 20 x 50mL

Test carried out according to current Pharmacopoeias.

**GLP**

Compliant  
Laboratory

*Suz Wood*

Date *18 Oct 2010*

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

**Sample Booking Number:** CN00154509

Certificate of Analysis - OSMM

Page 1 of 1

Flexible Medical Packaging Limited

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*