

Contact address:

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

## CERTIFICATE OF ANALYSIS

Date: *11/10/10*

<b>Product Name</b>	CLEARKLENS TEGO 2000SC VH25
<b>Product Code</b>	7513387
<b>Batch Number</b>	<i>W028519</i>
<b>Production Date</b>	<i>09/2010</i>
<b>Expiry Date</b>	<i>09/2012</i>

This is to certify that the above batch of product has been produced conforming to the manufacturing dossier from JohnsonDiversey in which the formula and processing instructions are defined and has been tested with the following Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	<i>clear slightly yellow liquid</i>
pH 100% (20°C)	JMD001	7.5 - 8.5	<i>7.6</i>
Specific Gravity (20°C)	JMD004	0.993 - 1.003	<i>1.001</i>

Signature / name *C. F. B. J. O. E.*  
On behalf of Manufacturing Location



FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P015928  
 Date Rec'd 23/09/10  
 Date: 27/09/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210090669	DIVERSEY TEGO 2000SC 50ml 25-42kGy	100	BATCH NO: W028519-CUST B/NO:327583  ACTUAL DOSE....31.8 - 33.9....kGy
B2FLE0010999 B210090669	SAMPLES	6	CUSTOMERS BATCH/CODE NO'S AS ORDER NO PO 15928 DATED 23.09.10
Despatch Text:  2PLTS			
Total		106	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Gaffey*

For and on behalf of ISOTRON LIMITED