



Contact

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 19/11/10

<b>Product Name</b>	CLEARKLENS TEGO 2001SC VH24
<b>Product Code</b>	7513402
<b>Batch Number</b>	WO 28641
<b>Production Date</b>	11/2010
<b>Expiry Date</b>	11/2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	8.5 - 9.5	8.9
Specific Gravity (20°C)	JDM004	1.060 - 1.070	1.061

"SIGNATURE" *e. Pascoe*

Name *Caroline Pascoe*  
Function Name *Pharmaceutical Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P016143  
 Date Rec'd 11/11/10  
 Date: 18/11/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B210110475	DIVERSEY TEG02001 10x5x100 25-42kGy	204	101xB/No.W028674 CUSTOMERB/No.349549 103xB/No.W028641 CUSTOMERB/No.347520  ACTUAL DOSE....35.8 - 36.5....kGy
B2FLE0010999 B210110475	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P016143 DATED 10/11/10
	Despatch Text:  4PLTS		
	Total	206	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED