

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 29/11/10

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	NO 28727
Production Date	11/2010
Expiry Date	11/2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	Clear Slightly yellow liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.6
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.997

"SIGNATURE" *C. Pascol*

Name *Caroline Pascol*
Function Name *Pharmaroom SUPERVISOR*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P016164
 Date Rec'd 18/11/10
 Date: 25/11/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B210110796	DIVERSEY TEGO 2000SC 50ml 25-42kGy	104	B/No.W028727-CUSTOMER B/No.330571 ACTUAL DOSE....35.3 - 35.9....kGy
B2FLE0010999 B210110796	SAMPLES	1	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P016164 DATED 17/11/10
	Despatch Text: 4PLTS		
	Total	105	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED