



**Contact**

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 17/12/10

<b>Product Name</b>	CLEARKLENS TEGO 2001SC VH24
<b>Product Code</b>	7513402
<b>Batch Number</b>	W028916
<b>Production Date</b>	12/2010
<b>Expiry Date</b>	12/2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	8.5 - 9.5	8.6
Specific Gravity (20°C)	JDM004	1.060 - 1.070	1.061

"SIGNATURE" C Pascoe

Name Caroline Pascoe  
Function Name Pharmaroom Supervisor

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P016290  
 Date Rec'd 9/12/10  
 Date: 13/12/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B210120354	DIVERSEY TEG02001 10x5x100 25-42kGy	186	B/No.W028916 - CUSTOMER B/No.342917  ACTUAL DOSE....34.5....kGy
B2FLE0010999 B210120354	SAMPLES	5	CUSTOMERS BATCH/CODE No.s AS ORDER No.P016290 DATED 8/12/10
Despatch Text:  4PLTS			
Total		191	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *T. Goffey*

For and on behalf of ISOTRON LIMITED