

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 10/11/10

Product Name	CLEARKLENS DE STANDARD VH29
Product Code	7515499
Batch Number	0028919
Production Date	11 / 2010
Expiry Date	11 / 2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.885

"SIGNATURE"

Geminushe

Name

Function Name *Pharma Supervisor*

Please note Johnsondiversev did change the company name to Diversey and is in the phase of implementation

CERTIFICATE OF IRRADIATION

Number: **B**

0470302 00

7 DEC 2010

FLE001

Page 1

Flexible Medical Packaging Ltd
Unit 8
White Cross Industrial Estate
Hightown
Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

LA1 4XS

P016223

25/11/10

6/12/10

Cust. Ref:

Date Rec'd

Date:

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210120042	JD DE 900ML BOTTLES 25-45kGy	355	B/No.W028919 - CUSTOMER B/No.2109 ACTUAL DOSE...35.6 - 37.3...kGy
P2FLE0010999 L10120042	SAMPLES	1	CUSTOMERS BATCH/CODE No.s AS ORDER No.P016223 DATED 24/11/10
Despatch Text: 7PLTS		Total	356 Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED