



Contact

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 13/12/10

<b>Product Name</b>	CLEARKLENS CLEANSINALD SC VH9
<b>Product Code</b>	7516429
<b>Batch Number</b>	W029048
<b>Production Date</b>	12/2010
<b>Expiry Date</b>	12/2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	12.0 - 13.0	12.1
Specific Gravity (20°C)	JDM004	1.040 - 1.060	1.045

"SIGNATURE" *C Pascoe*

Name *Caroline Pascoe*  
Function Name *Pharmaroom Supervisor*

Please note Johnsondiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 P016237

Cust. Ref:  
 Date Rec'd: 2/12/10  
 Date: 9/12/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE001013+ B210120218	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy             Despatch Text:  2PLTS	11	B/No. W029047 W029048  CUSTOMERS BATCH/CODE No.s AS ORDER No. P016237 DATED 2/12/10
Total		11	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 16 Dec 10  
**Date Tested:** 24 Dec 10  
**Date Completed:** 07 Jan 11  
**Purchase Order:** PO16320

**Test Article:** Clearklens Cleansinald SC  
Block 1  
**Sample Code:** VLF 99C01 10-103  
**Batch Ref:** WO29048  
**Qty Received:** 20x25mL  
**Laboratory Reference:** 0000848/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM06
Day Book Reference	MM107/04	000075
Volume Tested	MM107/05	20 x 25 mL

### External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 17 Jan 11 13:42:00

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

**Consignment:** 0000848      Certificate of Analysis - OSMM      Print Number      P0002430  
Lancaster

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*