

Contact

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 28/2/11

<b>Product Name</b>	CLEARKLENS TEGO 2001SC VH24
<b>Product Code</b>	7517177
<b>Batch Number</b>	W029142
<b>Production Date</b>	02/2011
<b>Expiry Date</b>	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	8.5 - 9.5	8.6
Specific Gravity (20°C)	JDM004	1.040 - 1.080	1.053

"SIGNATURE" *C. Pascoe*

Name *Caroline Pascoe*  
Function Name *Pharmaceutical Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 P016582

Cust. Ref: P016582  
 Date Rec'd 17/02/11  
 Date: 21/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B211020602	DIVERSEY TE602001 10x5x100 25-42kGy	202	105 X B/No..W029142-CUST B/No..356686 97 X B/No..W029328-CUST B/No..358415  ACTUAL DOSE....32.1 - 32.9....kGy
B2FLE0010999 B211020602	SAMPLES	2	CUSTOMERS BATCH/CODE No.s AS ORDER No.P016582 DATED 16/2/11
Despatch Text:  4FLTS			
Total		204	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 28 Feb 11  
**Date Tested:** 09 Mar 11  
**Date Completed:** 23 Mar 11  
**Purchase Order:** P016619

**Test Article:** Clearklens Tego 2001 SC  
VH24  
**Sample Code:** 356686  
**Batch Ref:** WO29142  
**Qty Received:** 10 x 100 mL Bottles ##  
**Laboratory Reference:** 0001179/2  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM03
Volume Tested	MM107/05	20 x 50 mL

**External Comments**

## 10 more bottles sent over on 03/03/11, totalling 20 x 100 mL Bottles

See Investigation Report PHM OOS 2011 018

**External Comments**

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 24 Mar 11 13:50:50

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Laboratory Manager - Pharmaceutical Microbiology

**Consignment:** 0001179      Certificate of Analysis - OSMM      Print Number      P0003044  
Lancaster

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*