



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 4.01.11

Product Name	CLEARKLENS TEGO 2001SC VH24
Product Code	7513402
Batch Number	NO 29289
Production Date	12 2010
Expiry Date	12 2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH (100% 20°C)	JDM001	8.5 - 9.5	8.5
Specific Gravity (20°C)	JDM004	1.060 - 1.070	1.060

"SIGNATURE" *C Pascoe*

Name *Caroline Pascoe*

Function Name *Pharmaceutical Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 3
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P016358
 Date Rec'd 22/12/10
 Date: 31/12/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B210120654	DIVERSEY TEG02001 10x5x100 25-42kGy	21	7XB/No.W029289 CUSTOMER B/No.353267 14XB/No.W029249 CUSTOMER B/No.349549 ACTUAL DOSE....32.0....kGy
B2FLE0010999 B210120654	SAMPLES	2	1XB/No.W029289 CUSTOMER B/No.353267 1XB/No.W029249 CUSTOMER B/No.349549 CUSTOMERS BATCH/CODE No.s AS ORDER No.P016358 DATED 22/12/10
Despatch Text:			
2PLTS			
Total		23	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffey*

For and on behalf of ISOTRON LIMITED