



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 21/2/11

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	W029323
Production Date	02/2011
Expiry Date	02/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.5
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.997

"SIGNATURE" *C Pascoe*

Name *Caroline Pascoe*
Function Name *Pharmaroom Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 P016547
 10/02/11
 14/02/11

Cust. Ref:
 Date Rec'd
 Date:

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B211020377	DIVERSEY TEGO 2000SC 50ml 25-42kGy	105	60XB/No.W029326 CUSTOMER B/No.350080 45XB/No.W029323 CUSTOMER B/No.340944 ACTUAL DOSE...32.4 - 33.0...kGy
B2FLE0010999 B211020377	SAMPLES	3	CUSTOMERS BATCH/CODE No.s AS ORDER No.P016547 DATED 9/2/11
B2FLE001MIN	MINIMUM CHARGE UPLIFT		
	Despatch Text: 3PLTS		
	Total	108	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED