

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 4. 01. 11

Product Name	CLEARKLENS DE STANDARD VH29
Product Code	7515499
Batch Number	NO 29333
Production Date	12 2010
Expiry Date	12 2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	0.885

"SIGNATURE" *C Pascoe*

Name *Caroline Pascoe*
Function Name *Pharmacy Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 Cust. Ref: P016358
 Date Rec'd 22/12/10
 Date: 31/12/10

Cust. Ref: P016358
 Date Rec'd 22/12/10
 Date: 31/12/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B210120657	JD DE 900ML BOTTLES 25-45kGy	174	B/No.W029333 ACTUAL DOSE....37.0 - 38.9....kGy
B2FLE0010999 B210120657	SAMPLES	1	 CUSTOMERS BATCH/CODE No.s AS ORDER No.P016358 DATED 22/12/10
	Despatch Text: 4PLTS		
	Total	175	Last Page 1 of 1

All in accordance with current technical agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED