



Contact

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 18/4/11

<b>Product Name</b>	CLEARKLENS TEGO 2000SC VH25
<b>Product Code</b>	7513387
<b>Batch Number</b>	W029759
<b>Production Date</b>	04/2011
<b>Expiry Date</b>	04/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	clear slightly yellow liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.5
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.997

"SIGNATURE" *C Pascoe*

Name *Caroline Pascoe*  
Function Name *Plamareom Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS  
 P016814  
 7/04/11  
 11/04/11

Cust. Ref: P016814  
 Date Rec'd 7/04/11  
 Date: 11/04/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B211040338	DIVERSEY TEGO 2000SC 50ml 25-42kGy	206	CUSTOMER B/No.363873 99 X B/No.W029758 106 X B/No.W029759  ACTUAL DOSE....32.8 - 34.4....kGy
B2FLE0010999 B211040338	SAMPLES	2	CUSTOMER BATCH/CODE No.s AS ORDER No.P016814 DATED 6/4/11
	Despatch Text:  4PLTS		
	Total	208	Last Page 1 of 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED