

**Contact**

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 27.05.2011

<b>Product Name</b>	CLEARKLENS CLEANSINALD SC VH9
<b>Product Code</b>	7516429
<b>Batch Number</b>	No 30099
<b>Production Date</b>	05/2011
<b>Expiry Date</b>	05/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	Clear Slightly Yellow Liquid
pH (100% 20°C)	JDM001	12.0 - 13.0	12.1
Specific Gravity (20°C)	JDM004	1.040 - 1.060	1.040

"SIGNATURE" *C. Perce*

Name *Cathrine Perce*

Function Name *Pharmacy Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Flexible Medical Packaging Ltd  
Unit 8  
White Cross Industrial Estate  
Hightown  
Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

LA1 4XS  
F016981

Cust. Ref: F016981  
Date Rec'd 19/05/11  
Date: 24/05/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010134 B211050600	C/SINALD BLK BOTT/CAPS/BAGS 25-40kGy	11	B/No.W030098 - W030099 CUSTOMER BATCH/CODE No.s AS ORDER No.P016981 DATED 18/5/11
Despatch Text: 2FLTS			
Total		11	Last Page 1 of 1

All in accordance with current Technical Agreement

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



## CERTIFICATE OF ANALYSIS

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

Date Received: 28 May 11  
Date Tested: 08 Jun 11  
Date Completed: 22 Jun 11  
Purchase Order: 17027

Test Article: ClearKlens Cleansinald SC  
Block 1  
Sample Code: VLF 99c 0110 340  
Batch Ref: WO30099  
Qty Received: 20 x 25mL  
Laboratory Reference: 0001694/1  
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM06
Volume Tested	MM107/05	20 x 25 mL

### External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 23 Jun 11 16:05:53

Mrs P. Pham-Lengoc, BSc, MSB  
Deputy Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0001694  
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004040

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*