



Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 28/11/11

Product Name	CLEARKLENS TEGO 2001SS VH24S
Product Code	7513403
Batch Number	W030275
Production Date	07/2011
Expiry Date	05/2012

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Liquid	clear liquid
pH (100% 20°C)	JDM001	8.5 - 9.2	8.7
Specific Gravity (20°C)	JDM004	0.998 - 1.001	1.001

"SIGNATURE" *C. Pascoe*

Name *Caroline Pascoe*

Function Name *Pharmacy Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P017226
 Date Rec'd 14/07/11
 Date: 26/07/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010093 B211070760	TEGO 2001 SS 25-45kGy	126	B/No.W030275 - CUSTOMER B/No.358415 ACTUAL DOSE .36.1 - 37.0 .kGy
B2FLE0010999 B211070760	SAMPLES	4	CUSTOMER BATCH/CODE No.s AS ORDER No.P017226 DATED 13/7/11
B2FLE001VAL	VALIDATION		TRIPPLICATE VALIDATION Ref:4.0968
Despatch Text: 3PLTS			
Total		130	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 01 Aug 11
Date Tested: 05 Aug 11
Date Completed: 19 Aug 11
Purchase Order: 17295

Test Article: Clarklens Tego 2001 SS
Block 1
Sample Code: 358415
Batch Ref: WO30275
Qty Received: 20 x 900mL
Laboratory Reference: 0002088/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM08
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Susan Wood

Date: 24 Aug 11 15:08:26

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002088
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004794

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited